



KVK/ 813/2025

Date: 06.01.2026

### QUOTATION NOTICE

Sealed quotations are invited for the Double Distillation Unit with the following specifications. The item has to be supplied and installed at the station. Interested firms are requested to participate in the quotation process.

#### SPECIFICATIONS:

- Double Distillation Unit with Distillation Apparatus Auto Cut off System – Borosilicate boiler, borosilicate condenser, quartz heater
- Distillation water output capacity - approx. 1.5 lit/h
- Minimum cooling water requirement - 1 lit/min
- Power consumption 1.5 kw
- Distillation temperature - 65-75°C
- Conductivity 3x10 s/cm.
- Suitable distillation apparatus power supply
- One-year full warranty

#### TERMS AND CONDITIONS

1. Sealed cover containing the quotation should be superscribed "Double Distillation Unit" and addressed to the 'Senior Scientist and Head, Krishi Vigyan Kendra - Kottayam, Kumarakom, and reach the undersigned on or before 10.30 am on 16.01.2026 and the same will be opened at 11:00 pm on the same day at the chamber of the head of office in the presence of those quotationers or their representatives (with letter of authority from the concerned quotationers) who are present at that time.
2. Late and incomplete quotations will not be considered at any circumstances.
3. The rate quoted should be inclusive of all taxes, transportation, loading & unloading, installation & fittings and other charges if any.
4. Warranty should be clearly specified.
5. The firm should have service personal in Kerala preferably at Kottayam. The firm should be nationally or internationally reputed and should be ISI/CE Certified.
6. The lowest rate offered as per the specifications will be accepted if reasonable.
7. Negotiations if any, will be conducted with the person/firm who has offered the lowest price
8. The item should be supplied within 7 days on receipt of supply order issued.

9. The rules and regulations pertaining to Government quotations will be applicable to this quotation.
10. The items should be delivered at the expenses of the successful bidder.
11. The firms/companies should ensure onsite service support for the items.
12. All parts of the items should be covered under the warranty. During warranty period, a minimum of 2 services has to be done.
13. The Quotationers should mention the email ID, GST and telephone/mobile number of the firm on the Quotation.
14. Senior Scientist and Head, KVK Kottayam shall have full power to accept or reject the quotation without assigning any reason, to do so.



Senior Scientist & Head

To: Notice Board, KAU website

